

# 11-3834

Pump Seal, PE

Comparable to: Agilent® 0905-1719



## Technical Data Sheet

---

### 1.0 Description

This seal is designed to be compatible with Agilent® 1290 pumps. It is comparable to Agilent® part number 0905-1719.

### 2.0 Materials Specifications

UHMW-PE

Elgiloy®

### 3.0 Replacement Frequency

Beyond routine maintenance it may be necessary to replace the seals more frequently if they are operating in a high salt buffer concentration environment; or if the plungers are scratched or worn. Seals may also need replacing if the system fails its leak test.

### 4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.

*Disclaimer: Sciencix, Inc. is not affiliated with or an authorized distributor/official partner of Original Equipment Manufacturer (OEM) parts. All OEM part names, numbers, brands, registered trademarks, or trademarks are the sole property of their owners in the US or any other country. All OEM part names, numbers, and company names used in this report are for cross reference purposes only. Use of these names and brands does not imply endorsement. Statements made by Sciencix regarding the quality and performance of our products are our own and should not be construed as an indication of OEM acceptance or approval.*