

# 11-2605

Needle Seat, Std., PEEK™, 0.17mm ID  
Comparable to: Agilent® G1329-87017



## Technical Data Sheet

---

### 1.0 Description

This needle seat assembly is designed to be compatible with Agilent® 1100, 1120, 1200, 1220, 1260, G1313A and G1329A/B HPLC systems. It is comparable to Agilent® part number G1329-87017.

### 2.0 Wetted Materials Specifications


316 SS  
PEEK™

### 3.0 Replacement Frequency

Replacement may be indicated by leaks at the needle seat or the accumulation of buffer at the top of the needle seat (indicating a small leak), or a blockage of the seat capillary is determined.

### 4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.