

CTS-10839

Seal Pack Rebuild Kit
Comparable to: Waters®
WAT271019



Technical Data Sheet

1.0 Description

This kit contains all the parts necessary to keep Waters® 2690, 2690D, 2695, 2695D and Alliance® HPLC systems working at peak performance. It is comparable to Waters® part number WAT271019.

2.0 Specifications

2.1 Components

Qty	Part Description
1	Wash Tube Seal (CTS-10598)
1	Tube, Seal Wash (CTS-10739)
1	PTFE Washer (CTS-10741-01)
1	Filter Retainer (CTS-10785)
1	Lower Wash Seal Frit (CTS-10788)
1	Needle Wash Frit (CTS-2324)
1	TFE Washer (CTS-7321)
1	Needle Assembly (CTS-7550)
2	Gold Injector Seal (CTS-A047)
1	Ferrule, SS (CTS-2330)
1	Compression Screw, SS (CTS-5313)

2.2 Materials

302 SS
316 SS
Delrin®
MP35N SS
PTFE
UHMWPE



3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.

4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.