

# CTS-3209

Syringe Seal, 250 $\mu$ L/2500 $\mu$ L  
Comparable to: Waters®  
WAT077347



## Technical Data Sheet

---

### 1.0 Description


This syringe seal is designed to be compatible with Waters® 717, 717 Plus, 2690, 2690D, 2695, 2695D, 2790, 2795 and Alliance® HPLC systems. It is comparable to Waters® part number WAT077347.

### 2.0 Materials Specifications

PTFE

### 3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.