

CTS-7162

Plunger Seal, Black
Compared to: Beckman® 237162, 728770



Technical Data Sheet

1.0 Description

This seal is designed to be compatible with Beckman® 114M, 116, 118, 125, 126, 127 and 128 HPLC systems. It is compared to Beckman® part numbers 237162 and 728770.

2.0 Materials Specifications

Carbon Filled PTFE
300 SS

3.0 Replacement Frequency

Beyond routine maintenance it may be necessary to replace the seals more frequently if they are operating in a high salt buffer concentration environment; or if the plungers are scratched or worn. Seals may also need replacing if the system fails its leak test.

4.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



5.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.