

CTS-11283

PM Kit for Standard Autosamplers
Comparable to: Agilent®
G1313-68730



Technical Data Sheet

1.0 Description

This kit contains all the parts necessary to keep Agilent® 1100, 1200, G1313, G1313A, G1327A and G1329A HPLC systems working at peak performance. It is comparable to Agilent® part number G1313-68730.

2.0 Specifications

2.1 Components

Qty	Part Description
1	Rotor Seal Assy (CTS-10459)
1	Needle Assembly (CTS-10835)
1	Needle Seat, Std., PEEK™, 0.17mm ID (11-2605)

2.2 Wetted Materials

Vespel®
304L SS
316Ti SS
PEEK™

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.