

# CTS-11288

Isolation Seal for 0101-0921 Valve  
Comparable to: Agilent®  
0100-1852



## Technical Data Sheet

---

### 1.0 Description

This isolation seal is designed to be compatible with Agilent® 1100 and 1200 HPLC systems. It is comparable to Agilent® part number 0100-1852.

### 2.0 Materials Specifications

Carbon Filled PTFE  
SS Spring

### 3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at [sales@sciencix.com](mailto:sales@sciencix.com) to report the issue. We will assist you with a resolution in a timely fashion.