# CTS-5746

Reference Valve Rebuild Kit Comparable to: Waters® WAT025746





## **Technical Data Sheet**

#### 1.0 Description

This rebuild kit is designed to be compatible with Waters 510, 515, 600, 610, 1515, 1525, 1525 Micro and 1525EF HPLC systems. It is comparable to Waters **previous version** of part number WAT025746.

### 2.0 Wetted Materials Specifications

440C SS Gold PTFE



Replacement may be indicated by leaks at the needle seat or the accumulation of buffer at the top of the needle seat (indicating a small leak), or a blockage of the seat capillary is determined.



#### 4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at <a href="mailto:sales@sciencix.com">sales@sciencix.com</a> to report the issue. We will assist you with a resolution in a timely fashion.