

CTS-10722

Autosampler Preventative Maintenance Kit
Comparable to: Agilent® G1313-68709



Technical Data Sheet

1.0 Description

This PM kit contains all the parts necessary to keep Agilent® 1100, 1200, G1313, G1313A, G1327A and G1329A HPLC systems working at peak performance. It is comparable to Agilent® part number G1313-68709.

2.0 Specifications

2.1 Components

Qty	Part Description
1	Rotor Seal Assy (CTS-10459)
1	Black Piston Seals, 2/pk (CTS-10417)
1	Needle Assembly (CTS-10835)
1	Needle Seat, Std., PEEK™, 0.17mm ID (11-2605)
1	Finger Caps, 3/pk (CTS-10834)

2.2 Wetted Materials

Carbon Filled PTFE
316 SS
Vespel®
PEEK™

3.0 Installation Information

Sciencix parts are designed to be compatible and interchangeable with OEM parts. Please see instrument service manual for proper removal and installation procedure.



4.0 Warranty

We view your business with us as a partnership based on mutual trust.

We extend a Lifetime Warranty against any manufacturing defects on all Sciencix HPLC repair parts. If a defect is identified, please contact us at 800.682.6480 or by email at sales@sciencix.com to report the issue. We will assist you with a resolution in a timely fashion.